

### INSTRUMENT CALIBRATION LIST

INSTRUMENT	MANUFACTURER	MODEL / SERIAL #	CALIBRATION DATE	CALIBRATION DUE DATE
Module Sensor	Evergreen Telemetry	MS-T&H-101 / 2300169C	3/27/2024	3/27/2025
Pressure Module	Evergreen Telemetry	S-PVF-1 / 2300389C	3/27/2024	3/27/2025
Humidity Sensor	Evergreen Telemetry	PR-TH-12 / 2300120	3/27/2024	3/27/2025
Immersion Temp Probe	Evergreen Telemetry	PR-T-4-6 / 2300180	3/27/2024	3/27/2025
Temp Probe	Evergreen Telemetry	PR-T-5 / 2300241	5/31/2023	5/31/2025
Water Meter	Dywer	490W-6-HKIT / 08Q2MW	12/11/2023	12/11/2024
AC/DC Clamp-on	Amprobe	AMP-220 / 170201139	1/2/2024	1/2/2025
Tachometer	Extech	461895 / H308432	1/2/2024	1/2/2025
Anemometer	Amprobe	TMA-10A / 12040330	1/2/2024	1/2/2025
Stroboscope / Photo-Tach	Extech	461825 / H461825	9/22/2023	9/22/2024



# Certificate of Calibration

Airflow Pros

Manufacturer	Evergreen Telemetry	Calibration Environment		
Temperature Product	Module Sensor	Temperature	74	°F
Model	MS - T&H - 101	Rel. Humidity	31	%
SN	2300169C	Bar. Pressure	28.8	in Hg

As Found     
  As Left     
  In Tolerance     
  Out of Tolerance

## Calibration Data

Measurement Variable	Test Point	Cal Standard	Allowable Range		Test Instrument	
			Min	Max		
Cal Lab Probe & Test Module	Spec					
Temperature (°F)	1	73.2	-0.3	+0.3	73.2	
	2	242.2	-2.6	+2.6	242.7	
	3	-44.0	-1.6	+1.6	-44.0	

Indicates out of tolerance condition -----↑

## Calibration Standard SN & Dates

Variable	System ID	Calibration Last	Calibration Due
Temperature	16320239	12-Sep-23	12-Sep-25
Temperature	21396189	5-Feb-24	5-Feb-26

This instrument has been checked for accuracy, calibrated to manufacturer's specifications, and found to be within the specified tolerance unless otherwise stated. It has been calibrated using measurement standards traceable to the National Institute of Standards and Technology, or accepted intrinsic standards of measurement, or derived by the ratio type of self calibrated techniques.

Calibrated By
 

 27-Mar-2024  
 Calibration Date
 

 27-Mar-2026  
 Date Due



# Certificate Of Calibration

Airflow Pros

Manufacturer	Evergreen Telemetry	Calibration Environment		
Product	Pressure / Velocity Module	Temperature	74	°F
Model	S-PVF-1	Rel. Humidity	31	%
SN	2300389C	Bar. Pressure	28.8	in Hg

As Found     
  As Left     
  In Tolerance     
  Out of Tolerance

## Calibration Data

Measurement Variable	Test Point	Cal Standard	Allowable Range		Test Instrument	
			Min	Max		
Barometric Pressure (in Hg)	Spec		-2% - 0.1	+ 2% + 0.1		
	1	20.0			20.0	
	2	28.9			29.0	
	3	33.0			33.1	
Differential Pressure (in wc)	Spec		-2%-.001	+2%+.001		
	1	10.00			9.971	
	2	2.000			1.992	
	3	0.5000			0.4965	
	4	0.0500			0.0495	
	5	-10.00			-10.016	
Via Pitot >> Velocity Pressure >> (inW.C. / FPM) -3% -7			-3% - 7	+3% + 7		
	7	0.00066 / 103			102	
	8	0.0159 / 505			504	

Indicates out of tolerance condition -----↑

## NIST-Traceable Lab Calibration Standards

Variable	System ID	Calibration Last	Calibration Due
Pressure	7481227	8-Mar-23	8-Mar-25
Pressure	7568470	8-Mar-23	8-Mar-25
Pressure	7871917	12-Sep-23	12-Sep-25
Pressure	7870754	12-Sep-23	12-Sep-25
Pressure	2205000006	13-Sep-23	13-Sep-25
Pressure	1208000080	13-Feb-23	13-Feb-25
Pressure	41001F6C	27-Apr-23	27-Apr-25
Velocity	2100191A	24-Feb-23	24-Feb-25
Velocity	2100190A	1-May-23	1-May-25

This instrument has been checked for accuracy, calibrated to manufacturer's specifications, and found to be within the specified tolerance unless otherwise stated. It has been calibrated using measurement standards traceable to the National Institute of Standards and Technology, or accepted intrinsic standards of measurement, or derived by the ratio type of self calibrated techniques.

\_\_\_\_\_  
 Calibrated By

27-Mar-2024      27-Mar-2026  
 Calibration Date      Date Due



# Certificate of Calibration

Airflow Pros

Manufacturer	Evergreen Telemetry	Calibration Environment		
Product	Humidity Sensor	Temperature	74	°F
Model	PR-TH-12	Rel. Humidity	31	%
SN	2300120	Bar. Pressure	28.8	in Hg

As Found     
  As Left     
  In Tolerance     
  Out of Tolerance

## Calibration Data

Measurement Variable	Test Point	Cal Standard	Allowable Range		Test Instrument
			Min	Max	
Temperature (°F)	Spec				
	1	40.1	-1.0	1.0	39.9
	2	75.2	-1.0	1.0	75.1
	3	85.9	-1.0	1.0	85.8
	4	127.5	-2.0	2.0	127.6
Barometric Pressure (in Hg)	Spec		-2% - 0.1	+ 2% + 0.1	
	1	20.0			19.9
	2	28.9			28.9
	3	33.0			33.0
Humidity %RH 10 to 90%	Spec		-3	3	
	1	9.5			10.0
	2	25.2			26.4
	3	58.4			56.3
	4	88.7			88.0

Indicates out of tolerance condition -----↑

## Calibration Standard

Variable	System ID	Calibration Last	Calibration Due
Temperature	16320239	12-Sep-23	12-Sep-25
Temperature	21396189	5-Feb-24	5-Feb-26
Pressure	2205000006	13-Sep-23	13-Sep-25
Pressure	1208000080	13-Feb-23	13-Feb-25
Humidity	20558772	12-Sep-23	12-Sep-24
Humidity	20052171	5-Feb-24	5-Feb-25

This instrument has been checked for accuracy, calibrated to manufacturer's specifications, and found to be within the specified tolerance unless otherwise stated. It has been calibrated using measurement standards traceable to the National Institute of Standards and Technology, or accepted intrinsic standards of measurement, or derived by the ratio type of self calibrated techniques.

Temperature accuracy (dry bulb) varies across the operating range:

Temperature over 32-100F	+/- 1.0 F
Temperature over 100-158F	+/- 2.0 F

---

 Calibrated By

27-Mar-2024      27-Mar-2025  


---

 Calibration Date      Date Due



# Certificate of Calibration

Airflow Pros

Manufacturer	Evergreen Telemetry		Calibration Environment		
Temperature Product	Module	Probe	Temperature	74	°F
Model		PR-T-4-6	Rel. Humidity	31	%
SN		2300180	Bar. Pressure	28.8	in Hg

As Found     
  As Left     
  In Tolerance     
  Out of Tolerance

## Calibration Data

Measurement Variable	Test Point	Cal Standard	Allowable Range		Test Instrument	
			Min	Max		
Cal Lab Module & Test Probe	Spec					
Temperature (°F)	1	73.3	-0.3	+0.3	73.3	
	2	242.3	-2.6	+2.6	242.1	
	3	-43.9	-1.6	+1.6	-43.8	

Indicates out of tolerance condition -----↑

## Calibration Standard SN & Dates

Variable	System ID	Calibration Last	Calibration Due
Temperature	16320239	12-Sep-23	12-Sep-25
Temperature	21396189	5-Feb-24	5-Feb-26

This instrument has been checked for accuracy, calibrated to manufacturer's specifications, and found to be within the specified tolerance unless otherwise stated. It has been calibrated using measurement standards traceable to the National Institute of Standards and Technology, or accepted intrinsic standards of measurement, or derived by the ratio type of self calibrated techniques.

\_\_\_\_\_  
 Calibrated By

27-Mar-2024      27-Mar-2026  
 Calibration Date      Date Due



# Certificate of Calibration

Airflow Pros

Manufacturer	Evergreen Telemetry			Calibration Environment		
Temperature Product	Module	Probe	Temperature	75	°F	
Model		PR-T-5	Rel. Humidity	24	%	
SN		2300241	Bar. Pressure	28.5	in Hg	

As Found     
  As Left     
  In Tolerance     
  Out of Tolerance

## Calibration Data

Measurement Variable	Test Point	Cal Standard	Allowable Range		Test Instrument	
			Min	Max		
Cal Lab Module & Test Probe	Spec					
Temperature (°F)	1	78.4	-0.3	+0.3	78.3	
	2	242.3	-2.6	+2.6	243.2	
	3	-43.1	-1.6	+1.6	-43.0	

Indicates out of tolerance condition ----↑

## Calibration Standard SN & Dates

Variable	System ID	Calibration Last	Calibration Due
Temperature	21396189	5-Oct-21	5-Oct-23

This instrument has been checked for accuracy, calibrated to manufacturer's specifications, and found to be within the specified tolerance unless otherwise stated. It has been calibrated using measurement standards traceable to the National Institute of Standards and Technology, or accepted intrinsic standards of measurement, or derived by the ratio type of self calibrated techniques.

\_\_\_\_\_
   
 Calibrated By

31-May-2023      31-May-2025  
 Calibration Date      Date Due



## CERTIFICATE OF NIST CALIBRATION

Dwyer Instruments, 102 Highway 212, Michigan City, IN 46360  
 USA T: +1 800.872.9141 +1 219.879.8000 F: +1 219.872.9057

ID/Serial#: 08Q2MW	Date: 12/11/2023	Date Due: 12/11/2024	By: 3968																
<b>Customer Information</b>		<b>Device Under Test (DUT)</b>																	
AIRFLOW PROS 1001 EASTWIN DRIVE SUITE 203 WESTERVILLE, OH 43081	ID/Serial#: 08Q2MW Model: 490W-6-HKIT Description: Hydronic Differential Pressure Manometer MFR: Dwyer Instruments Accuracy: 2% Of Reading Red Sensor: DWY-H200-08Q2MW Device ID Blue Sensor: DWY-L200-08Q2MW Device ID																		
<b>Address Where Calibration Was Performed</b>		<b>Calibration Standard Information</b>																	
Dwyer Instruments Inc 3999 Hupp Rd Kingsbury, IN 46345	Instrument Reference: Status: <table style="display: inline-table; border: none;"><tr><td style="border: none;">As Received</td><td style="border: none;">After Repair</td><td style="border: none;">New</td></tr><tr><td style="border: none; text-align: center;"><input type="checkbox"/></td><td style="border: none; text-align: center;"><input type="checkbox"/></td><td style="border: none; text-align: center;"><input checked="" type="checkbox"/></td></tr></table>	As Received	After Repair	New	<input type="checkbox"/>	<input type="checkbox"/>	<input checked="" type="checkbox"/>												
As Received		After Repair	New																
<input type="checkbox"/>	<input type="checkbox"/>	<input checked="" type="checkbox"/>																	
<b>Reference Standards Used</b>	Test Range: 0 - 200 PSI																		
<table border="1" style="width: 100%; border-collapse: collapse;"> <thead> <tr> <th style="width: 15%;">Module</th> <th style="width: 15%;">ID#</th> <th style="width: 15%;">Last Cal</th> <th style="width: 15%;">Due Date</th> </tr> </thead> <tbody> <tr> <td>MW-MFC015</td> <td>22DWYPSL-1296</td> <td>3/17/23</td> <td>3/17/24</td> </tr> <tr> <td>MW-PM293</td> <td>22DWYPSL-1183</td> <td>12/14/22</td> <td>12/14/23</td> </tr> <tr> <td>MW-PM294</td> <td>22DWYPSL-1245</td> <td>1/3/23</td> <td>1/3/24</td> </tr> </tbody> </table>	Module	ID#	Last Cal	Due Date	MW-MFC015	22DWYPSL-1296	3/17/23	3/17/24	MW-PM293	22DWYPSL-1183	12/14/22	12/14/23	MW-PM294	22DWYPSL-1245	1/3/23	1/3/24	Notes:		
Module	ID#	Last Cal	Due Date																
MW-MFC015	22DWYPSL-1296	3/17/23	3/17/24																
MW-PM293	22DWYPSL-1183	12/14/22	12/14/23																
MW-PM294	22DWYPSL-1245	1/3/23	1/3/24																
Master Gauge Accuracy: 0.05%		Certificate No: 23DWY09-1350 Sales order No: S932879 Purchase order No: CREDITCARD																	

As Found/As Left Calibration Data - All Data on PSI			
RESULT	Target Test Point	Reference Standard Reading	Error Percentage
PASS	-0.015	-0.030	0.0074
PASS	120.115	119.940	0.0874
PASS	199.850	199.940	-0.0450
PASS	0.052	-0.010	0.0312
PASS	99.088	99.340	-0.1260
PASS	49.231	49.440	-0.1046
PASS	40.105	39.990	0.0574
PASS	80.143	79.960	0.0913
PASS	160.022	159.940	0.0408
PASS	149.028	149.220	-0.0962

This document certifies that only the device under test (DUT) identified above has been calibrated against a reference standard having an accuracy as listed.

Calibrated By: ML AL

Job Function: Technician

Hoover Instrument Service, Inc.  
 401 North Home Road  
 Mansfield, Ohio 44906-2398  
 (419) 529-3226 Fax(419) 529-9360

To: **Airflow Pros**  
**1001 Eastwind Drive - Suite 203**  
**Westerville, OH 43081**

Test Report # 63612  
 P.O.# \_\_\_\_\_  
 Date: 1/2/24

Condition as Received  Returned  In Tolerance  Out of Tolerance  Data on Next Sheet

Make:	Amprobe	Hoover		Customer		Hoover		Customer	
		DCV	ACV	Resistance	DCV	ACV	Resistance		
Description:	ac/dc clamp-on	50.00	49.9	50.00	49.9	100.0	100.0	100.0	100.0
Model #	AMP-220	100.0	99.8	100.0	99.8	1.000 K	1.000 K	1.000 K	1.000 K
Mfg Serial #	170201139	150.0	149.8	150.0	149.8	10.000 K	9.99K		
Customer #		250.0	249.7	250.0	249.7				
RH %	36	500.0	499.6	500.0	499.6				
Temperature:	21 °C								
Accuracy:	acv = +/- 1% +/- 5d								
	aca = +/- 1.8% +/- 5d								
	dcv = +/- 1% +/- 5d								
	dca = +/- 2% +/- 5d								
	res = +/- 1% +/- 5d								
Accuracy of Standard:									
	acv = +/- 0.5%								
	aca = +/- 0.5%								
	dcv = +/- 0.03%								
	dca = +/- 0.3%								
	res. = +/- 0.1%								

Services Performed:

- Cleaned       Movement Repaired       Pointer Repaired       Reset Pointer  
 Zero Adjusted       Adjust Calibration       Repaired       Batteries

Calibration Location On Site

Test Instruments Used:

**Amprobe** #BDM40-UA s/n 07110052  
**Shallcross** #830 s/n 21107

Due Date:

6/24  
4/24

Notes:

Calibration Procedure: **5075-33**

Revision: **0**

Calibration Cycle: **1 year**

Tested By:

*[Signature]*

Quality Control:

*[Signature]*



Hoover Instrument Service, Inc.  
 401 North Home Road  
 Mansfield, Ohio 44906-2398  
 (419) 529-3226 Fax(419) 529-9360

To: **Airflow Pros**  
**1001 Eastwind Drive - Suite 203**  
**Westerville, Ohio 43081**

Test Report # 63618  
 P.O.# \_\_\_\_\_  
 Date: 1/2/24

Condition as Received  Returned  In Tolerance  Out of Tolerance  Data on Next Sheet

	Hoover	Customer	Hoover	Customer	Hoover	Customer
			<b>RPM</b>			
<b>Make:</b> <u>Extech</u>						
<b>Description:</b> <u>tachometer</u>			600.0	599.9		
<b>Model #</b> <u>461895</u>			900.0	899.9		
<b>Mfg Serial #</b> <u>H308432</u>			1800.0	1800		
<b>Customer #</b> _____			3000.0	3000		
<b>RH %</b> <u>36</u> <b>Temperature:</b> <u>21</u> °C			3600.0	3600		
<b>Accuracy:</b> +/- 0.05% +/- 1d			6000.0	6000		
			7200.0	7200		
<b>Accuracy of Standard:</b> +/- 0.0008%						

**Services Performed:**

- Cleaned
- Movement Repaired
- Pointer Repaired
- Reset Pointer
- Zero Adjusted
- Adjust Calibration
- Repaired
- Batteries
- Calibration Location On Site

**Test Instruments Used:**

**Altek** #40A s/n 195587

**Due Date:**

8/24

**Notes:**

**Calibration Procedure:** 5023

**Revision:** 4

**Calibration Cycle:** 1 year

**Tested By:** [Signature]

**Quality Control:** [Signature]

Hoover Instrument Service, Inc.  
 401 North Home Road  
 Mansfield, Ohio 44906-2398  
 (419) 529-3226 Fax(419) 529-9360

To: **Airflow Pros**  
**1001 Eastwind Drive - Suite 203**  
**Westerville, OH 43081**

Test Report # 63615  
 P.O.# \_\_\_\_\_  
 Date: 1/2/24

Condition as Received  Returned  In Tolerance  Out of Tolerance  Data on Next Sheet

	Hoover		Customer		Hoover		Customer	
					feet/minute		temperature fahrenheit	
Make:	<b>Amprobe</b>							
Description:	<b>anemometer</b>				<u>260</u>	<u>265</u>	<u>69.5</u>	<u>69.4</u>
Model #	<b>TMA-10A</b>				<u>510</u>	<u>522</u>		
Mfg Serial #	<b>12040330</b>				<u>740</u>	<u>750</u>		
Customer #					<u>1010</u>	<u>1018</u>		
RH %	<u>36</u>	Temperature:	<u>21</u>	°C	<u>1260</u>	<u>1265</u>		
Accuracy:	<b>Temp = +/- 1.5 °F</b>				<u>1500</u>	<u>1510</u>		
	<b>+/- 2% of F.S.</b>							
Accuracy of Standard:	<b>+/- 5 ft/min</b>							
	<b>Temp = +/- 0.5 °F</b>							

Services Performed:

- Cleaned
- Movement Repaired
- Pointer Repaired
- Reset Pointer
- Zero Adjusted
- Adjust Calibration
- Repaired
- Batteries

Calibration Location On Site

Test Instruments Used:

<b>Tegam</b>	#855	s/n T-113227
<b>TSI</b>	#8330	s/n 95030100

Due Date:

12/24  
6/24


Notes:

Replaced Cable

Calibration Procedure: **5075-23**

Revision: **0**

Calibration Cycle: **1 year**

Tested By: 

Quality Control: 

Hoover Instrument Service, Inc.  
 401 North Home Road  
 Mansfield, Ohio 44906-2398  
 (419) 529-3226 Fax(419) 529-9360

To: **Airflow Pros**  
**1001 Eastwind Drive - Suite 203**  
**Westerville, Ohio 43081**

Test Report # 63213  
 P.O.# \_\_\_\_\_  
 Date: 9/22/23

Condition as Received  Returned  In Tolerance  Out of Tolerance  Data on Next Sheet

Make:	Hoover	Customer	Hoover		Customer	
			Photo-Tach	RPM	Stroboscope	RPM
<b>Extech</b>						
<b>Description: stroboscope / photo-tach</b>						
<b>Model # 461825</b>			300.0	300.0	300.0	299.9
<b>Mfg Serial # H437768</b>			600.0	600.0	600.0	600.0
<b>Customer #</b>			900.0	900.0	900.0	899.9
<b>RH % 54 Temperature: 23 °C</b>			1800	1800	1800	1800
<b>Accuracy: +/- 0.1 RPM +/- 2d</b>			3600	3600	3600	3600
<b>Accuracy of Standard:</b>			7200	7200	7200	7200
<b>strobe = +/- 0.05%</b>						
<b>photo = +/- 0.0008%</b>						

Services Performed:

- Cleaned
- Movement Repaired
- Pointer Repaired
- Reset Pointer
- Zero Adjusted
- Adjust Calibration
- Repaired
- Batteries

Calibration Location On Site

Test Instruments Used:

**Altek** #40A s/n 195587  
**Sperry** #TACH-1 s/n L301594

Due Date:

8/24  
4/24

Notes:

Calibration Procedure: **5023-2**

Revision: **1**

Calibration Cycle: **1 year**

Tested By:

Quality Control: