

INSTRUMENT CALIBRATION LIST

INSTRUMENT	MANUFACTURER	MODEL / SERIAL #	CALIBRATION DATE	CALIBRATION DUE DATE
Immersion Temp	Evergreen	PR-T-4-6 / 2200098	6/17/2024	6/17/2025
Water Meter	Dwyer	490W-6 / 08Q2KP	6/20/2024	6/20/2025
AC/DC Clamp-on	AMPROBE	AMP-220 / 200501518	6/14/2024	6/14/2025
Stroboscope / Photo-Tach	Extech	461825 / H437762	6/14/2024	6/14/2025
Module Sensor	Evergreen Telemetry	MS-T&H-101 / 2100150B	6/17/2024	6/17/2025
Humidity Sensor	Evergreen Telemetry	PR-TH-12 / 2300111	6/18/2024	6/18/2025
Humidity Sensor	Evergreen Telemetry	PR-TH-1 / 2200180	6/18/2024	6/18/2025
Pressure Module	Evergreen Telemetry	S-PVF-1 / 2200373C	6/17/2024	6/17/2025



Certificate of Calibration

Airflow Pros

Manufacturer	Evergreen Telemetry		Calibration Environment			
Temperature Product		Module	Probe	Temperature	71	°F
Model			PR-T-4-6	Rel. Humidity	26	%
SN			2200098	Bar. Pressure	28.3	in Hg

As Found
 As Left
 In Tolerance
 Out of Tolerance

Calibration Data

Measurement Variable	Test Point	Cal Standard	Allowable Range		Test Instrument
			Min	Max	
Cal Lab Module & Test Probe	Spec				
Temperature (°F)	1	75.0	-0.3	+0.3	74.9
	2	242.4	-2.6	+2.6	241.8
	3	-43.8	-1.6	+1.6	-43.4

Indicates out of tolerance condition -----↑

Calibration Standard SN & Dates

Variable	System ID	Calibration Last	Calibration Due
Temperature	16320239	12-Sep-23	12-Sep-25
Temperature	21396189	5-Feb-24	5-Feb-26

This instrument has been checked for accuracy, calibrated to manufacturer's specifications, and found to be within the specified tolerance unless otherwise stated. It has been calibrated using measurement standards traceable to the National Institute of Standards and Technology, or accepted intrinsic standards of measurement, or derived by the ratio type of self calibrated techniques.

 Calibrated By

17-Jun-2024 17-Jun-2026

 Calibration Date Date Due



CERTIFICATE OF NIST CALIBRATION

Dwyer Instruments, 102 Highway 212, Michigan City, IN 46360
 USA T: +1 800.872.9141 +1 219.879.8000 F: +1
 219.872.9057

ID/Serial#: 08Q2KP		Date: 6/20/2024	Date Due: 09/19/2024	By: 3968
Customer Information		Device Under Test (DUT)		
Airflow Pros 4741 Atlantic Blvd, Ste. B2 000044248 Jacksonville, FL 32207		ID/Serial#:	08Q2KP	
		Model:	490W-6-NIST	
		Description:	Hydronic Differential Pressure Manometer	
		MFR:	Dwyer Instruments	
		Accuracy:	2% Of Reading	
		Red Sensor:	DWY-H200-08Q2KP	Device ID
		Blue Sensor:	DWY-L200-08Q2KP	Device ID
Address Where Calibration Was Performed		Calibration Standard Information		
Dwyer Instruments Inc 3999 Hupp Rd Kingsbury, IN 46345		Instrument Reference:		
		Status:	As Received <input checked="" type="checkbox"/>	After Repair <input type="checkbox"/>
			New <input type="checkbox"/>	
Reference Standards Used		Test Range:	0 - 200 PSI	
Module	ID#	Last Cal	Due Date	
MW-MFC052	24DWYPSL-0468	5/9/24	8/9/24	
MW-PM294	24DWYPSL-0027	1/10/24	1/10/25	
MW-PM288	24DWYPSL-0104	2/8/24	8/8/24	
Master Gauge Accuracy:		0.05%		
		Notes:	Certificate No: 24DWY09-754	
			Sales order No: S002795	
			Purchase order No: 000044248	

As Found/As Left Calibration Data - All Data on PSI			
RESULT	Target Test Point	Reference Standard Reading	Error Percentage
PASS	-0.009	-0.010	0.0003
PASS	119.226	119.960	-0.3668
PASS	198.392	199.960	-0.7838
PASS	2.041	0.020	1.0107
PASS	99.377	99.450	-0.0364
PASS	50.342	49.440	0.4508
PASS	39.798	40.000	-0.1011
PASS	79.542	79.980	-0.2188
PASS	158.845	159.960	-0.5576
PASS	148.499	149.350	-0.4256

This document certifies that only the device under test (DUT) identified above has been calibrated against a reference standard having an accuracy as listed.

Calibrated By: Bonnie N. Smith Job Function: Technician

Hoover Instrument Service, Inc.
 401 North Home Road
 Mansfield, Ohio 44906-2398
 (419) 529-3226 Fax(419) 529-9360

To: **Airflow Pros**
 4741 Atlantic Blvd - Suite 2B
 Jacksonville, FL 32207

Test Report # 64728
 P.O.# _____
 Date: 6/14/24

Condition as Received Returned In Tolerance Out of Tolerance Data on Next Sheet

Make:	Amprobe	Hoover		Customer		Hoover		Customer	
		DCV		ACV		Resistance			
Description:	ac/dc clamp-on	50.00	49.8	50.00	49.8	100.0	100.0	1.000 K	0.999K
Model #	AMP-220	100.0	99.9	100.0	99.8	10.000 K	9.99K		
Mfg Serial #	200501518	150.0	149.8	150.0	149.7				
		250.0	249.7	250.0	249.3				
		500.0	499.6	500.0	499.0				
Customer #									
RH %	55	DCA		ACA					
Temperature:	22 °C	25.0	25.05	25.0	25.22				
Accuracy:	acv = +/- 1% +/- 5d	50.0	50.33	50.0	50.24				
	aca = +/- 1.8% +/- 5d	75.0	75.4	75.0	75.4				
	dcv = +/- 1% +/- 5d	100.0	100.6	100.0	100.7				
	dca = +/- 2% +/- 5d	200.0	200.6	200.0	202.0				
	res = +/- 1% +/- 5d	300.0	302.6	300.0	304.2				
Accuracy of Standard:		400.0	403.5	400.0	405.8				
	acv = +/- 0.5% dcv = +/- 0.03%	500.0	504.5	500.0	506.5				
	aca = +/- 0.5% dca = +/- 0.3%								
	res = +/- 0.1%								

Services Performed:

- Cleaned Movement Repaired Pointer Repaired Reset Pointer
 Zero Adjusted Adjust Calibration Repaired Batteries
 Calibration Location On Site

Test Instruments Used:

Amprobe	#BDM40-UA	s/n 07110052
Shallcross	#830	s/n 21107

Due Date:

12/24
4/25

Notes:

Calibration Procedure: 5075-33

Revision: 0

Calibration Cycle: 1 year

Tested By:

[Signature]

Quality Control:

[Signature]

Hoover Instrument Service, Inc.
 401 North Home Road
 Mansfield, Ohio 44906-2398
 (419) 529-3226 Fax(419) 529-9360

To: **Airflow Pros**
 4741 Atlantic Blvd - Suite 2B
 Jacksonville, FL 32207

Test Report # 64729
 P.O.# _____
 Date: 6/14/24

Condition as Received Returned In Tolerance Out of Tolerance Data on Next Sheet

	Hoover		Customer		Hoover		Customer	
Make:	Extech				RPM		RPM	
Description:	stroboscope / photo-tach				Photo-Tach		Stroboscope	
Model #	461825				300.0	<u>300.0</u>	300.0	<u>300.0</u>
Mfg Serial #	H437762				600.0	<u>599.9</u>	600.0	<u>600.0</u>
Customer #					900.0	<u>899.9</u>	900.0	<u>899.9</u>
RH %	<u>55</u>	Temperature:	<u>22</u> °C		1800	<u>1800</u>	1800	<u>1800</u>
Accuracy:					3600	<u>3600</u>	3600	<u>3599</u>
	<u>+/- 0.1 RPM +/- 2d</u>				7200	<u>7200</u>	7200	<u>7200</u>
Accuracy of Standard:								
	<u>strobe = +/- 0.05%</u>							
	<u>photo = +/- 0.0008%</u>							

- Services Performed:
- Cleaned
 - Movement Repaired
 - Pointer Repaired
 - Reset Pointer
 - Zero Adjusted
 - Adjust Calibration
 - Repaired
 - Batteries
 - Calibration Location On Site

Test Instruments Used:			Due Date:	Notes:
Altek	#40A	s/n 195587	<u>8/24</u>	
Sperry	#TACH-1	s/n L301594	<u>4/25</u>	

Calibration Procedure: 5023-2 Revision: 1
 Calibration Cycle: 1 year

Tested By: [Signature] Quality Control: [Signature]



Certificate of Calibration

Airflow Pros

Manufacturer	Evergreen Telemetry	Calibration Environment		
Temperature Product	Module Sensor	Temperature	71	°F
Model	MS - T&H - 101	Rel. Humidity	26	%
SN	2100150B	Bar. Pressure	28.3	In Hg

As Found
 As Left
 In Tolerance
 Out of Tolerance

Calibration Data

Measurement Variable	Test Point	Cal Standard	Allowable Range		Test Instrument
			Min	Max	
Cal Lab Probe & Test Module	Spec				
Temperature (°F)	1	75.5	-0.3	+0.3	75.5
	2	242.5	-2.6	+2.6	241.3
	3	-43.6	-1.6	+1.6	-42.5

Indicates out of tolerance condition -----↑

Calibration Standard SN & Dates

Variable	System ID	Calibration Last	Calibration Due
Temperature	16320239	12-Sep-23	12-Sep-25
Temperature	21396189	5-Feb-24	5-Feb-26

This instrument has been checked for accuracy, calibrated to manufacturer's specifications, and found to be within the specified tolerance unless otherwise stated. It has been calibrated using measurement standards traceable to the National Institute of Standards and Technology, or accepted intrinsic standards of measurement, or derived by the ratio type of self calibrated techniques.

 Calibrated By

17-Jun-2024 17-Jun-2026

 Calibration Date Date Due



Certificate of Calibration

Airflow Pros

Manufacturer	Evergreen Telemetry	Calibration Environment		
Product	Humidity Sensor	Temperature	70	°F
Model	PR-TH-12	Rel. Humidity	30	%
SN	2300111	Bar. Pressure	28.4	in Hg

As Found
 As Left
 In Tolerance
 Out of Tolerance

Calibration Data

Measurement Variable	Test Point	Cal Standard	Allowable Range		Test Instrument	
			Min	Max		
Temperature (°F)	Spec					
	1	39.9	-1.0	1.0	39.9	
	2	75.1	-1.0	1.0	75.0	
	3	85.8	-1.0	1.0	85.4	
	4	127.3	-2.0	2.0	127.2	
Barometric Pressure (in Hg)	Spec		-2% - 0.1	+ 2% + 0.1		
	1	20.0			20.0	
	2	28.4			28.4	
	3	33.0			33.0	
	4					
Humidity %RH 10 to 90%	Spec		-3	3		
	1	9.9			10.1	
	2	22.6			23.7	
	3	61.1			59.4	
	4	90.1			89.2	

Indicates out of tolerance condition -----↑

Calibration Standard

Variable	System ID	Calibration Last	Calibration Due
Temperature	16320239	12-Sep-23	12-Sep-25
Temperature	21396189	5-Feb-24	5-Feb-26
Pressure	2205000006	13-Sep-23	13-Sep-25
Pressure	1208000080	13-Feb-23	13-Feb-25
Humidity	20558772	12-Sep-23	12-Sep-24
Humidity	20052171	5-Feb-24	5-Feb-25

This instrument has been checked for accuracy, calibrated to manufacturer's specifications, and found to be within the specified tolerance unless otherwise stated. It has been calibrated using measurement standards traceable to the National Institute of Standards and Technology, or accepted intrinsic standards of measurement, or derived by the ratio type of self calibrated techniques.

Temperature accuracy (dry bulb) varies across the operating range:

Temperature over 32-100F +/- 1.0 F
 Temperature over 100-158F +/- 2.0 F

 Calibrated By

18-Jun-2024 18-Jun-2025
 Calibration Date Date Due



Certificate of Calibration

Airflow Pros

Manufacturer	Evergreen Telemetry	Calibration Environment		
Product	Humidity Sensor	Temperature	70	°F
Model	PR-TH-1	Rel. Humidity	30	%
SN	2200180	Bar. Pressure	28.4	in Hg

As Found
 As Left
 In Tolerance
 Out of Tolerance

Calibration Data

Measurement Variable	Test Point	Cal Standard	Allowable Range		Test Instrument
			Min	Max	
Temperature (°F)	Spec				
	1	39.9	-1.0	1.0	40.2
	2	75.1	-1.0	1.0	75.0
	3	85.8	-1.0	1.0	85.4
	4	127.2	-2.0	2.0	126.5
Barometric Pressure (in Hg)	Spec		-2% - 0.1	+ 2% + 0.1	
	1	20.0			20.0
	2	28.4			28.4
	3	33.0			33.0
	4				
Humidity %RH 10 to 90%	Spec		-3	3	
	1	9.9			10.5
	2	23.1			24.2
	3	61.2			58.6
	4	90.4			89.8

Indicates out of tolerance condition -----↑

Calibration Standard

Variable	System ID	Calibration Last	Calibration Due
Temperature	16320239	12-Sep-23	12-Sep-25
Temperature	21396189	5-Feb-24	5-Feb-26
Pressure	2205000006	13-Sep-23	13-Sep-25
Pressure	1208000080	13-Feb-23	13-Feb-25
Humidity	20558772	12-Sep-23	12-Sep-24
Humidity	20052171	5-Feb-24	5-Feb-25

This instrument has been checked for accuracy, calibrated to manufacturer's specifications, and found to be within the specified tolerance unless otherwise stated. It has been calibrated using measurement standards traceable to the National Institute of Standards and Technology, or accepted intrinsic standards of measurement, or derived by the ratio type of self calibrated techniques.

Temperature accuracy (dry bulb) varies across the operating range:

Temperature over 32-100F +/- 1.0 F
 Temperature over 100-158F +/- 2.0 F

Calibrated By

18-Jun-2024 18-Jun-2025
 Calibration Date Date Due



Certificate Of Calibration

Airflow Pros

Manufacturer	Evergreen Telemetry	Calibration Environment		
Product	Pressure / Velocity Module	Temperature	71	°F
Model	S-PVF-1	Rel. Humidity	26	%
SN	2200373C	Bar. Pressure	28.3	in Hg

As Found
 As Left
 In Tolerance
 Out of Tolerance

Calibration Data

Measurement Variable	Test Point	Cal Standard	Allowable Range		Test Instrument
			Min	Max	
Barometric Pressure (in Hg)	Spec		-2% - 0.1	+ 2% + 0.1	
	1	20.0			20.1
	2	28.3			28.4
	3	33.0			33.1
Differential Pressure (in wc)	Spec		-2%-0.001	+2%+.001	
	1	10.00			9.974
	2	2.000			1.992
	3	0.5000			0.4960
	4	0.0500			0.0496
	5	-10.00			-9.997
	6	-0.0500			-0.0498
Via Pitot >>	7	0.00065 / 102	-3% - 7	+3% + 7	102
Velocity Pressure >>	8	0.0158 / 503			501
(inW.C. / FPM)					
-3% -7					

Indicates out of tolerance condition -----↑

NIST-Traceable Lab Calibration Standards

Variable	System ID	Calibration Last	Calibration Due
Pressure	7481227	8-Mar-23	8-Mar-25
Pressure	7568470	8-Mar-23	8-Mar-25
Pressure	7871917	12-Sep-23	12-Sep-25
Pressure	7870754	12-Sep-23	12-Sep-25
Pressure	2205000006	13-Sep-23	13-Sep-25
Pressure	1208000080	13-Feb-23	13-Feb-25
Pressure	41001F6C	27-Apr-23	27-Apr-25
Velocity	2100191A	24-Feb-23	24-Feb-25
Velocity	2100190A	1-May-23	1-May-25

This instrument has been checked for accuracy, calibrated to manufacturer's specifications, and found to be within the specified tolerance unless otherwise stated. It has been calibrated using measurement standards traceable to the National Institute of Standards and Technology, or accepted intrinsic standards of measurement, or derived by the ratio type of self calibrated techniques.

 Calibrated By

17-Jun-2024 17-Jun-2026
 Calibration Date Date Due