



### INSTRUMENT CALIBRATION LIST

INSTRUMENT	MANUFACTURER	MODEL / SERIAL #	CALIBRATION DATE	CALIBRATION DUE DATE
Module Sensor	Evergreen Telemetry	T&H-101 / 2300230C	10/2/2024	10/2/2025
Water Meter	Alnor	HM-670 / 71201119	1/2/2024	1/2/2024
AC/DC Clamp-On	Amprobe	AMP-220 / 221200069	10/14/2024	10/14/2025
Anemometer	Amprobe	TMA-10A / 011323-1	10/14/2024	10/14/2025
Immersion Temp Probe	Evergreen Telemetry	PR-T-4-6 / 2300286	10/2/2024	10/2/2025
Humidity Sensor	Evergreen Telemetry	PR-TH-12 / 2300126	10/2/2024	10/2/2025
Pressure Module	Evergreen Telemetry	S-PVF-1 / 2100744C	10/1/2024	10/1/2025
Stroboscope/Photo-Tach	Extech	461825 / H443221	10/14/2024	10/14/2025



# Certificate of Calibration

Airflow Pros

Manufacturer	Evergreen Telemetry	Calibration Environment		
Temperature Product	Module Sensor	Temperature	70	°F
Model	MS - T&H - 101	Rel. Humidity	35	%
SN	2300230C	Bar. Pressure	28.6	in Hg

As Found     
  As Left     
  In Tolerance     
  Out of Tolerance

## Calibration Data

Measurement Variable	Test Point	Cal Standard	Allowable Range		Test Instrument
			Min	Max	
Cal Lab Probe & Test Module	Spec				
Temperature (°F)	1	75.1	-0.3	+0.3	75.0
	2	242.8	-2.6	+2.6	241.8
	3	-43.4	-1.6	+1.6	-43.5

Indicates out of tolerance condition ----↑

## Calibration Standard SN & Dates

Variable	System ID	Calibration Last	Calibration Due
Temperature	16320239	12-Sep-23	12-Sep-25
Temperature	21396189	5-Feb-24	5-Feb-26

This instrument has been checked for accuracy, calibrated to manufacturer's specifications, and found to be within the specified tolerance unless otherwise stated. It has been calibrated using measurement standards traceable to the National Institute of Standards and Technology, or accepted intrinsic standards of measurement, or derived by the ratio type of self calibrated techniques.

\_\_\_\_\_  
 Calibrated By

2-Oct-2024      2-Oct-2026  
 Calibration Date      Date Due

Hoover Instrument Service, Inc.  
 401 North Home Road  
 Mansfield, Ohio 44906-2398  
 (419) 529-3226 Fax(419) 529-9360

To: **Airflow Pros**  
 1001 Eastwind Drive - Suite 203  
 Westerville, Ohio 43081

Test Report # 63611  
 P.O.# \_\_\_\_\_  
 Date: 1/2/24

Condition as Received  Returned  In Tolerance  Out of Tolerance  Data on Next Sheet

Make:	Alnor	Hoover		Customer		Hoover		Customer	
		inches of water	PSI	feet of water	PSI	feet of water	PSI	feet of water	
Description:	hydronic manometer	25.00	25.0	5.00	4.960	1.000	0.97		
		50.00	50.0			2.000	1.97		
Model #	HM-670	75.00	75.0	10.00	9.957	3.000	2.97		
		100.00	99.9			4.000	3.97		
Mfg Serial #	71201119	150.00	149.9	25.00	24.90	5.000	4.98		
		200.00	199.9			7.500	7.46		
Customer #		250.00	249.8	50.00	49.88	10.00	9.96		
		300.00	299.7			25.00	24.91		
RH %	36	350.00	349.6	75.00	74.83	50.00	49.81		
		400.00	399.5			75.00	74.74		
Accuracy:		500.00	499.2			100.00	99.62		
		750.00	748.7			150.00	149.5		
	+/- 1% of rdg.	1000.00	997.6						
	or +/- 0.036 psi	1250.00	1247						
	+/- 0.08 ft of water	1500.00	1497						
	+/- 1.0 in of water								
	whichever is greater								
Accuracy of Standard:									
	+/- 0.1%								

Services Performed:

- Cleaned       Movement Repaired       Pointer Repaired       Reset Pointer  
 Zero Adjusted       Adjust Calibration       Repaired       Batteries

Calibration Location On Site

Test Instruments Used:

Due Date:

Notes:

Crystal #300psiXP2i s/n 821664  
 Crystal #30psiXP2i s/n 462180

6/24  
 2/24

Calibration Procedure: 5040-33

Revision: 0

Calibration Cycle: 1 year

Tested By: [Signature]

Quality Control: [Signature]

Hoover Instrument Service, Inc.  
 401 North Home Road  
 Mansfield, Ohio 44906-2398  
 (419) 529-3226 Fax(419) 529-9360

To: **Airflow Pros**  
 1001 Eastwind Drive - Suite 203  
 Westerville, OH 43081

Test Report # 65321  
 P.O.# \_\_\_\_\_  
 Date: 10/14/24

Condition as Received  Returned  In Tolerance  Out of Tolerance  Data on Next Sheet

Make:	Amprobe	DCV		ACV		Resistance	
		Hoover	Customer	Hoover	Customer	Hoover	Customer
Description:	ac/dc clamp-on	50.00	49.9	50.00	49.8	100.0	100.0~
Model #	AMP-220	100.0	99.8	100.0	99.7	1.000 K	0.999K
Mfg Serial #	221200069	150.0	149.8	150.0	149.6	10.000 K	10.00K
		250.0	249.7	250.0	249.4		
		500.0	499.4	500.0	499.1		

Customer #	RH % <u>53</u> Temperature: <u>21</u> °C	DCA		ACA	
		Hoover	Customer	Hoover	Customer
		25.0	25.13	25.0	25.10
Accuracy:	acv = +/- 1% +/- 5d	50.0	50.30	50.0	50.35
	aca = +/- 1.8% +/- 5d	75.0	75.4	75.0	75.4
	dcv = +/- 1% +/- 5d	100.0	100.6	100.0	100.8
	dca = +/- 2% +/- 5d	200.0	201.4	200.0	202.2
	res = +/- 1% +/- 5d	300.0	301.3	300.0	302.0
		400.0	402.5	400.0	402.6
		500.0	503.5	500.0	503.8

Accuracy of Standard:  
 acv = +/- 0.5%      dcv = +/- 0.03%  
 aca = +/- 0.5%      dca = +/- 0.3%  
 res. = +/- 0.1%

- Services Performed:
- Cleaned
  - Zero Adjusted
  - Movement Repaired
  - Adjust Calibration
  - Pointer Repaired
  - Repaired
  - Reset Pointer
  - Batteries
  - Calibration Location On Site

Test Instruments Used:      Due Date:      Notes:

<b>Amprobe</b> #BDM40-UA      s/n 07110052	<u>12/24</u> <u>4/25</u>	
<b>Shallcross</b> #830      s/n 21107		

Calibration Procedure: 5075-33      Revision: 0  
 Calibration Cycle: 1 year

Tested By: [Signature]      Quality Control: [Signature]

Hoover Instrument Service, Inc.  
401 North Home Road  
Mansfield, Ohio 44906-2398  
(419) 529-3226 Fax(419) 529-9360

To: **Airflow Pros**  
**1001 Eastwind Drive - Suite 203**  
**Westerville, OH 43081**

Test Report # 65322  
P.O.# \_\_\_\_\_  
Date: 10/14/24

Condition as Received  Returned  In Tolerance  Out of Tolerance  Data on Next Sheet

	Hoover	Customer	Hoover	Customer	Hoover	Customer
Make:	<b>Amprobe</b>		<b>feet/minute</b>		<b>temperature fahrenheit</b>	
Description:	<b>anemometer</b>		395	391	69.4	70.4
Model #	<b>TMA-10A</b>		590	582		
Mfg Serial #	<b>011323-1</b>		820	810		
Customer #			1000	986		
RH % <u>53</u>	Temperature: <u>21</u> °C		1300	1290		
Accuracy:	Temp = +/- 1.5 °F +/- 2% of F.S.		1650	1636		
Accuracy of Standard:	+/- 5 ft/min Temp = +/- 0.5 °F					

Services Performed:

- Cleaned
- Movement Repaired
- Pointer Repaired
- Reset Pointer
- Zero Adjusted
- Adjust Calibration
- Repaired
- Batteries
- Calibration Location On Site

Test Instruments Used:

Tegam #855 s/n T-113227  
TSI #8330 s/n 95030100

Due Date:

12/24 -  
6/25

Notes:

Calibration Procedure: **5075-23**

Revision: **0**

Calibration Cycle: **1 year**

Tested By:

Quality Control:



# Certificate of Calibration

Airflow Pros

Manufacturer	Evergreen Telemetry			Calibration Environment		
Temperature Product		Module	Probe	Temperature	70	°F
Model			PR-T-4-6	Rel. Humidity	35	%
SN			2300286	Bar. Pressure	28.6	in Hg

As Found     
  As Left     
  In Tolerance     
  Out of Tolerance

## Calibration Data

Measurement Variable	Test Point	Cal Standard	Allowable Range		Test Instrument	
			Min	Max		
Cal Lab Module & Test Probe	Spec					
Temperature (°F)	1	75.1	-0.3	+0.3	75.0	
	2	242.8	-2.6	+2.6	241.8	
	3	-43.4	-1.6	+1.6	-43.5	

Indicates out of tolerance condition -----↑

## Calibration Standard SN & Dates

Variable	System ID	Calibration Last	Calibration Due
Temperature	16320239	12-Sep-23	12-Sep-25
Temperature	21396189	5-Feb-24	5-Feb-26

This instrument has been checked for accuracy, calibrated to manufacturer's specifications, and found to be within the specified tolerance unless otherwise stated. It has been calibrated using measurement standards traceable to the National Institute of Standards and Technology, or accepted intrinsic standards of measurement, or derived by the ratio type of self calibrated techniques.

\_\_\_\_\_  
 Calibrated By

2-Oct-2024      2-Oct-2026  
 Calibration Date      Date Due



# Certificate of Calibration

Airflow Pros

Manufacturer	Evergreen Telemetry	Calibration Environment		
Product	Humidity Sensor	Temperature	70	°F
Model	PR-TH-12	Rel. Humidity	35	%
SN	2300126	Bar. Pressure	28.6	in Hg

As Found     
  As Left     
  In Tolerance     
  Out of Tolerance

## Calibration Data

Measurement Variable	Test Point	Cal Standard	Allowable Range		Test Instrument	
			Min	Max		
Temperature (°F)	Spec					
	1	40.0	-1.0	1.0	40.2	
	2	75.1	-1.0	1.0	75.3	
	3	85.9	-1.0	1.0	86.0	
	4	128.6	-2.0	2.0	128.7	
Barometric Pressure (in Hg)	Spec		-2% - 0.1	+ 2% + 0.1		
	1	20.0			20.0	
	2	28.6			28.7	
	3	33.0			33.1	
Humidity %RH 10 to 90%	Spec		-3	3		
	1	10.1			9.8	
	2	23.8			25.2	
	3	59.0			59.0	
	4	90.3			88.8	

Indicates out of tolerance condition —↑

## Calibration Standard

Variable	System ID	Calibration Last	Calibration Due
Temperature	16320239	12-Sep-23	12-Sep-25
Temperature	21396189	5-Feb-24	5-Feb-26
Pressure	2205000006	13-Sep-23	13-Sep-25
Pressure	1208000080	13-Feb-23	13-Feb-25
Humidity	20558772	18-Sep-24	18-Sep-25
Humidity	20052171	5-Feb-24	5-Feb-25

This instrument has been checked for accuracy, calibrated to manufacturer's specifications, and found to be within the specified tolerance unless otherwise stated. It has been calibrated using measurement standards traceable to the National Institute of Standards and Technology, or accepted intrinsic standards of measurement, or derived by the ratio type of self calibrated techniques.

Temperature accuracy (dry bulb) varies across the operating range:

Temperature over 32-100F      +/- 1.0 F  
 Temperature over 100-158F      +/- 2.0 F

\_\_\_\_\_  
 Calibrated By

2-Oct-2024      2-Oct-2025  
 Calibration Date      Date Due





# Certificate Of Calibration

Airflow Pros

Manufacturer	Evergreen Telemetry	Calibration Environment		
Product	Pressure / Velocity Module	Temperature	75	°F
Model	S-PVF-1	Rel. Humidity	35	%
SN	2100744C	Bar. Pressure	28.6	in Hg

As Found     
  As Left     
  In Tolerance     
  Out of Tolerance

## Calibration Data

Measurement Variable	Test Point	Cal Standard	Allowable Range		Test Instrument
			Min	Max	
Barometric Pressure (in Hg)	Spec		-2% - 0.1	+ 2% + 0.1	
	1	20.0			20.1
	2	28.6			28.6
	3	33.0			33.1
Differential Pressure (in wc)	Spec		-2%-.001	+2%+.001	
	1	10.00			9.977
	2	2.000			1.994
	3	0.5000			0.0499
	4	0.0500			0.0504
	5	-10.00			-10.027
Via Pitot >> Velocity Pressure >> (inW.C. / FPM) -3% -7			-3% - 7	+3% + 7	
	7	0.00071 / 107			107
	8	0.0157 / 502			503

Indicates out of tolerance condition -----↑

## NIST-Traceable Lab Calibration Standards

Variable	System ID	Calibration Last	Calibration Due
Pressure	7481227	8-Mar-23	8-Mar-25
Pressure	7568470	8-Mar-23	8-Mar-25
Pressure	7871917	12-Sep-23	12-Sep-25
Pressure	7870754	12-Sep-23	12-Sep-25
Pressure	2205000006	13-Sep-23	13-Sep-25
Pressure	1208000080	13-Feb-23	13-Feb-25
Pressure	41001F6C	19-Jun-24	19-Jun-26
Velocity	2100191A	24-Feb-23	24-Feb-25
Velocity	2100190A	1-May-23	1-May-25

This instrument has been checked for accuracy, calibrated to manufacturer's specifications, and found to be within the specified tolerance unless otherwise stated. It has been calibrated using measurement standards traceable to the National Institute of Standards and Technology, or accepted intrinsic standards of measurement, or derived by the ratio type of self calibrated techniques.

\_\_\_\_\_  
 Calibrated By

1-Oct-2024      1-Oct-2026  
 \_\_\_\_\_  
 Calibration Date      Date Due



Hoover Instrument Service, Inc.  
 401 North Home Road  
 Mansfield, Ohio 44906-2398  
 (419) 529-3226 Fax(419) 529-9360

To: **Airflow Pros**  
**1001 Eastwind Drive - Suite 203**  
**Westerville, OH 43081**

Test Report # 65323  
 P.O.# \_\_\_\_\_  
 Date: 10/14/24

Condition as Received  Returned  In Tolerance  Out of Tolerance  Data on Next Sheet

Make:	Hoover	Customer	RPM		RPM	
			Hoover	Customer	Hoover	Customer
<b>Extech</b>			<b>Photo-Tach</b>		<b>Stroboscope</b>	
<b>Description:</b> <u>stroboscope / photo-tach</u>			300.0	<u>300.0</u>	300.0	<u>300.0</u>
<b>Model #</b> <u>461825</u>			600.0	<u>600.0</u>	600.0	<u>600.0</u>
<b>Mfg Serial #</b> <u>H443221</u>			900.0	<u>900.0</u>	900.0	<u>900.0</u>
<b>Customer #</b> _____			1800	<u>1800</u>	1800	<u>1800</u>
<b>RH %</b> <u>53</u> <b>Temperature:</b> <u>21 °C</u>			3600	<u>3600</u>	3600	<u>3600</u>
<b>Accuracy:</b> <u>+/- 0.1 RPM +/- 2d</u>			7200	<u>7200</u>	7200	<u>7201</u>

**Accuracy of Standard:**  
 strobe = +/- 0.05%  
 photo = +/- 0.0008%

**Services Performed:**

- Cleaned
- Movement Repaired
- Pointer Repaired
- Reset Pointer
- Zero Adjusted
- Adjust Calibration
- Repaired
- Batteries
- Calibration Location On Site

**Test Instruments Used:**

**Altek** #40A s/n 195587  
**Sperry** #TACH-1 s/n L301594

**Due Date:**

8/25  
4/25

**Notes:**

**Calibration Procedure:** 5023-2

**Revision:** 1

**Calibration Cycle:** 1 year

**Tested BY:**

*[Signature]*

**Quality Control:**

*[Signature]*